

ANALYTICAL REPORT

Job Number: 420-105379-1

Job Description: Sullivan County BOCES

For:

Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Attention: Jesse Morrill



Meredith W Ruthven
Customer Service Manager
mruthven@envirotestlaboratories.com
06/21/2016

NYSDOH ELAP does not certify for all parameters. EnviroTest Laboratories does hold certification for all analytes where certification is offered by ELAP unless otherwise specified. Pursuant to NELAP, this report may not be reproduced, except in full, without written approval of the laboratory. EnviroTest Laboratories Inc. certifies that the analytical results contained herein apply only to the samples tested as received by our laboratory. All questions regarding this report should be directed to the EnviroTest Customer Service Representative.

EnviroTest Laboratories, Inc. Certifications and Approvals: NYSDOH 10142, NJDEP NY015, CTDOH PH-0554

METHOD SUMMARY

Client: Sullivan County BOCES

Job Number: 420-105379-1

Description	Lab Location	Method	Preparation Method
Matrix: Water			
ICPMS Metals by 200.8	EnvTest	EPA 200.8 Rev.5.4	
200 Series Drinking Water Prep Determination Step	EnvTest		EPA 200

Lab References:

EnvTest = EnviroTest

Method References:

EPA = US Environmental Protection Agency

METHOD / ANALYST SUMMARY

Client: Sullivan County BOCES

Job Number: 420-105379-1

Method	Analyst	Analyst ID
EPA 200.8 Rev.5.4	Sirico, Derek	DS

SAMPLE SUMMARY

Client: Sullivan County BOCES

Job Number: 420-105379-1

<u>Lab Sample ID</u>	<u>Client Sample ID</u>	<u>Client Matrix</u>	<u>Date/Time Sampled</u>	<u>Date/Time Received</u>
420-105379-1	KIT-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-2	KIT-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-3	KIT-52-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-4	KIT-52-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-5	KIT-53-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-6	KIT-53-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-7	KIT-54-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-8	KIT-54-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-9	KIT-55-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-10	KIT-55-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-11	CAF-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-12	CAF-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-13	CAFB-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-14	CAFB-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-15	CAFG-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-16	CAFG-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-17	CAF-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-18	CAF-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-19	CAF-D2-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-20	CAF-D2-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-21	BAN-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-22	BAN-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-23	TL-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-24	TL-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-25	TEC-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-26	TEC-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-27	BG-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-28	BG-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-29	BG-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-30	BG-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-31	TEC-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-32	TEC-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-33	PEO-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-34	PEO-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-35	CUS-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-36	CUS-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-37	PO-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-38	PO-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-39	NO-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-40	NO-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-41	NO-52-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-42	NO-52-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-43	001-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-44	001-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-45	001T-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-46	001T-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020

SAMPLE SUMMARY

Client: Sullivan County BOCES

Job Number: 420-105379-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
420-105379-47	001-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-48	001-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-49	001G-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-50	001G-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-51	002-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-52	002-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-53	002T-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-54	002T-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-55	002-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-56	002-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-57	003-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-58	003-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-59	003T-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-60	003T-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-61	003-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-62	003-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-63	004B-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-64	004B-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-65	004-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-66	004-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-67	004T-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-68	004T-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-69	004-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-70	004-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-71	005-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-72	005-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-73	006-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-74	006-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-75	007-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-76	007-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-77	008-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-78	008-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-79	011-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-80	011-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-81	011-D1-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-82	011-D1-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-83	011G-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-84	011G-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-85	012B-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-86	012B-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-87	013-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-88	013-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-89	014-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-90	014-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-91	015-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-92	015-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020

SAMPLE SUMMARY

Client: Sullivan County BOCES

Job Number: 420-105379-1

Lab Sample ID	Client Sample ID	Client Matrix	Date/Time Sampled	Date/Time Received
420-105379-93	016-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-94	016-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-95	017-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-96	017-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-97	012-51-1	Drinking Water	06/07/2016 0900	06/07/2016 1020
420-105379-98	012-51-2	Drinking Water	06/07/2016 0900	06/07/2016 1020

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-51-1
Lab Sample ID: 420-105379-1

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1904		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	3.31	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-51-2
Lab Sample ID: 420-105379-2

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1907		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-52-1
Lab Sample ID: 420-105379-3

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1911		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.85	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-52-2
Lab Sample ID: 420-105379-4

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1914		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-53-1
Lab Sample ID: 420-105379-5

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1918		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.30	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-53-2
Lab Sample ID: 420-105379-6

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1921		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-54-1
Lab Sample ID: 420-105379-7

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1925		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	3.00	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-54-2
Lab Sample ID: 420-105379-8

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/10/2016 1928	
Prep Method: 200			Date Prepared:	06/10/2016 0930	
Pb	2.48	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-55-1
Lab Sample ID: 420-105379-9

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1932		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	2.33	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: KIT-55-2
Lab Sample ID: 420-105379-10

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1935		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-51-1
Lab Sample ID: 420-105379-11

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1946		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.83	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-51-2
Lab Sample ID: 420-105379-12

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1949		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.27	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAFB-51-1
Lab Sample ID: 420-105379-13

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1953		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAFB-51-2
Lab Sample ID: 420-105379-14

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 1956		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAFG-51-1
Lab Sample ID: 420-105379-15

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2000		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAFG-51-2
Lab Sample ID: 420-105379-16

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2004		
Prep Method: 200			Date Prepared: 06/10/2016 0930		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-D1-1
Lab Sample ID: 420-105379-17

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016	2007	
Prep Method: 200			Date Prepared: 06/10/2016	0930	
Pb	2.55	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-D1-2
Lab Sample ID: 420-105379-18

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2011		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-D2-1
Lab Sample ID: 420-105379-19

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2032		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.67	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CAF-D2-2
Lab Sample ID: 420-105379-20

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4				Date Analyzed: 06/10/2016 2035	
Prep Method: 200				Date Prepared: 06/10/2016 1030	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BAN-51-1
Lab Sample ID: 420-105379-21

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2039		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	2.72	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BAN-51-2
Lab Sample ID: 420-105379-22

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2042		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TL-51-1
Lab Sample ID: 420-105379-23

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2046		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.52	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TL-51-2
Lab Sample ID: 420-105379-24

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2049		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TEC-51-1
Lab Sample ID: 420-105379-25

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2053		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	9.12	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TEC-51-2
Lab Sample ID: 420-105379-26

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2056		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.46	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BG-51-1
Lab Sample ID: 420-105379-27

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2100		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	9.40	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BG-51-2
Lab Sample ID: 420-105379-28

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2111		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	2.66	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BG-D1-1
Lab Sample ID: 420-105379-29

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2114		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	4.69	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: BG-D1-2
Lab Sample ID: 420-105379-30

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2118		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.97	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TEC-D1-1
Lab Sample ID: 420-105379-31

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2121		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.50	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: TEC-D1-2
Lab Sample ID: 420-105379-32

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2125		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	4.29	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: PEO-51-1
Lab Sample ID: 420-105379-33

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2128		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	12.4	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: PEO-51-2
Lab Sample ID: 420-105379-34

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2132		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CUS-51-1
Lab Sample ID: 420-105379-35

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2135		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	2.81	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: CUS-51-2
Lab Sample ID: 420-105379-36

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2139		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: PO-51-1
Lab Sample ID: 420-105379-37

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2142		
Prep Method: 200			Date Prepared: 06/10/2016 1030		
Pb	1.98	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: PO-51-2
Lab Sample ID: 420-105379-38

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2153		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: NO-51-1
Lab Sample ID: 420-105379-39

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2207		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.43	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: NO-51-2
Lab Sample ID: 420-105379-40

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2211		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: NO-52-1
Lab Sample ID: 420-105379-41

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2214		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.67	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: NO-52-2
Lab Sample ID: 420-105379-42

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2218		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001-51-1
Lab Sample ID: 420-105379-43

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2221		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	3.18	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001-51-2
Lab Sample ID: 420-105379-44

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2225		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001T-51-1
Lab Sample ID: 420-105379-45

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2236		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	2.43	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001T-51-2
Lab Sample ID: 420-105379-46

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2239		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001-D1-1
Lab Sample ID: 420-105379-47

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2243		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	3.66	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001-D1-2
Lab Sample ID: 420-105379-48

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2246		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.04	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001G-51-1
Lab Sample ID: 420-105379-49

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2250		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.38	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 001G-51-2
Lab Sample ID: 420-105379-50

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2253		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002-51-1
Lab Sample ID: 420-105379-51

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2257		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	3.21	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002-51-2
Lab Sample ID: 420-105379-52

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2300		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002T-51-1
Lab Sample ID: 420-105379-53

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2304		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	2.76	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002T-51-2
Lab Sample ID: 420-105379-54

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2308		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002-D1-1
Lab Sample ID: 420-105379-55

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2318		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	12.5	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 002-D1-2
Lab Sample ID: 420-105379-56

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/10/2016 2322	
Prep Method: 200			Date Prepared:	06/10/2016 1130	
Pb	4.43	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003-51-1
Lab Sample ID: 420-105379-57

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2325		
Prep Method: 200			Date Prepared: 06/10/2016 1130		
Pb	3.00	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003-51-2
Lab Sample ID: 420-105379-58

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2329		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003T-51-1
Lab Sample ID: 420-105379-59

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4				06/10/2016 2343	
Prep Method: 200				06/10/2016 1330	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003T-51-2
Lab Sample ID: 420-105379-60

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2347		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.26	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003-D1-1
Lab Sample ID: 420-105379-61

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/10/2016 2350		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	4.52	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 003-D1-2
Lab Sample ID: 420-105379-62

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4				06/11/2016 0001	
Prep Method: 200				06/10/2016 1330	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004B-51-1
Lab Sample ID: 420-105379-63

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0004		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004B-51-2
Lab Sample ID: 420-105379-64

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0008		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004-51-1
Lab Sample ID: 420-105379-65

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0012		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.44	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004-51-2
Lab Sample ID: 420-105379-66

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0015	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	4.08	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004T-51-1
Lab Sample ID: 420-105379-67

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0019	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	1.44	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004T-51-2
Lab Sample ID: 420-105379-68

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0022		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004-D1-1
Lab Sample ID: 420-105379-69

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0026		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.79	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 004-D1-2
Lab Sample ID: 420-105379-70

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0029		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 005-51-1
Lab Sample ID: 420-105379-71

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0033		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	2.05	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 005-51-2
Lab Sample ID: 420-105379-72

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0044		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 006-51-1
Lab Sample ID: 420-105379-73

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0047		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	3.33	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 006-51-2
Lab Sample ID: 420-105379-74

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4					
Prep Method: 200					
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 007-51-1
Lab Sample ID: 420-105379-75

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0054	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	1.55	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 007-51-2
Lab Sample ID: 420-105379-76

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0058		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 008-51-1
Lab Sample ID: 420-105379-77

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0101		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 008-51-2
Lab Sample ID: 420-105379-78

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0105	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	3.33	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011-51-1
Lab Sample ID: 420-105379-79

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/15/2016 0516	
Prep Method: 200			Date Prepared:	06/13/2016 1400	
Pb	6.13	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011-51-2
Lab Sample ID: 420-105379-80

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/15/2016 0520	
Prep Method: 200			Date Prepared:	06/13/2016 1400	
Pb	1.31	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011-D1-1
Lab Sample ID: 420-105379-81

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0126		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011-D1-2
Lab Sample ID: 420-105379-82

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0130		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011G-51-1
Lab Sample ID: 420-105379-83

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0134		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	2.27	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 011G-51-2
Lab Sample ID: 420-105379-84

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0137		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 012B-51-1
Lab Sample ID: 420-105379-85

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0141		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	2.71	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 012B-51-2
Lab Sample ID: 420-105379-86

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0144		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 013-51-1
Lab Sample ID: 420-105379-87

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0148		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.22	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 013-51-2
Lab Sample ID: 420-105379-88

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0151		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 014-51-1
Lab Sample ID: 420-105379-89

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0155		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.85	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 014-51-2
Lab Sample ID: 420-105379-90

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0158	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 015-51-1
Lab Sample ID: 420-105379-91

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed:	06/11/2016 0209	
Prep Method: 200			Date Prepared:	06/10/2016 1330	
Pb	3.65	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 015-51-2
Lab Sample ID: 420-105379-92

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0213		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 016-51-1
Lab Sample ID: 420-105379-93

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0216		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.87	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 016-51-2
Lab Sample ID: 420-105379-94

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4					
Prep Method: 200					
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 017-51-1
Lab Sample ID: 420-105379-95

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0223		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	2.23	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 017-51-2
Lab Sample ID: 420-105379-96

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0227		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 012-51-1
Lab Sample ID: 420-105379-97

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0231		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	2.69	ug/L	1.00	1.00	1.0

Jesse Morrill
Sullivan County BOCES
6 Weirk Avenue
Liberty, NY 12754

Job Number: 420-105379-1

Client Sample ID: 012-51-2
Lab Sample ID: 420-105379-98

Date Sampled: 06/07/2016 0900
Date Received: 06/07/2016 1020
Client Matrix: Drinking Water

Analyte	Result/Qualifier	Unit	RL	RL	Dilution
Method: 200.8 Rev.5.4			Date Analyzed: 06/11/2016 0234		
Prep Method: 200			Date Prepared: 06/10/2016 1330		
Pb	1.00 U	ug/L	1.00	1.00	1.0

DATA REPORTING QUALIFIERS

Client: Sullivan County BOCES

Job Number:

Lab Section	Qualifier	Description
Metals	U	The analyte was analyzed for but not detected at or above the lowest stated limit.

Certification Information

Client: Sullivan County BOCES

Job Number:

The following analytes are Not Part of the ELAP scope of accreditation:

Sulfur, Tungsten, Silicon, Bicarbonate Alkalinity, 7 Day BOD 5210C, 28 Day BOD, Soluble BOD, Carbon Dioxide, carbonate Alkalinity, CBOD Soluble, Chlorine, Cyanide (WAD), Ferrous Iron, Ferric Iron, Total Nitrogen, Total Organic Nitrogen, Dissolved Oxygen, pH, Phenolphthalein Alkalinity, Solids (Fixed), Solids (Percent), Solids (Percent Moisture), Solids (Percent Volatile), Solids (Volatile Suspended), Temperature, TKN (Soluble), Total Inorganic Carbon, Volatile Acids as Acetic Acid, 2-Aminopyridine, 3-Picoline, 1-Methyl-2-pyrrolidinone, Aziridine, Dimethyl sulfoxide, Fluorobenzene, 1-Chlorohexane, Iron Bacteria, Salmonella, & Sulfur Reducing Bacteria.

The following analytes are Not Part of ELAP Potable Water scope of accreditation:

Cobalt (200.7, 200.8), Tin (200.7), Strontium (200.7), Gold (200.7), Platinum (200.7), Palladium (200.7), Titanium (200.7), Phosphorus (365.3), Nitrate-Nitrite (10-107-4-1C, 353.2), m-Xylene & p-Xylene (502.2, 524), Naphthalene (502.2), o-Xylene (502.2, 524), & Fecal Coliform (9222D).

The following analytes are Not Part of ELAP Solid and Hazardous Waste scope of accreditation:

Ammonia (SM 4500NH3G), Nitrate-Nitrite (353.2, 10-107-4-1C), TKN (351.2), Phosphorus (365.3), Total Cresols (8270), 1,2-Dichloro-1,1,2-trifluoroethane (8260), & Chlorodifluoromethane (8260).

The following analytes are Not Part of ELAP Non Potable Water scope of accreditation:

Dissolved Organic Carbon (5310C), Mecoprop (8151A), & MCPA (8151A).

The following analytes are Part of ELAP scope of accreditation but not for the noted methods:

Nitrate (Solid & Hazardous Waste Matrix, 300), Nitrite (Solid & Hazardous Waste, 300, 4500NO2), Sulfate (Solid & Hazardous, 300.0), alpha-Chlordane (608), Endrin Ketone (608), gamma-Chlordane (608), PCB-1262 (608), PCB-1268 (608), 1,2-Diphenylhydrazine (625), 2-MethylNaphthalene (625), 3-Methylphenol (625), 4-Nitroaniline (625), 1,1,1,2-Tetrachloroethane (624,601), 1,1,2-Trichloro-1,2,2-trifluoroethane (601,624), 1,2,3-Trichlorobenzene (624, 601), 1,2,3-Trichloropropane (624), 1,2,4-Trichlorobenzene (601,624), 1,2,4-trimethylbenzene (624), 1,2-Dichloro-3-Chloropropane (601, 624), 1,2-Dichloro-1,1,2-trifluoroethane (601, 624), 1,3,5-Trimethylbenzene (624), 1,3-Dichloropropane (624), 2,2-dichloropropane (601,624), 2-chlorotoluene (601,624), 2-hexanone (624), 4-Chlorotoluene (601,624), 4-Isopropyltoluene (624), Acetonitrile (624), Benzyl Chloride (624, 8021), Bromobenzene (601,624), Carbon disulfide (624), Bromochloromethane (624), Dibromomethane (624), 1,2-Dibromoethane (624), Hexachlorobutadiene (624), Isopropylbenzene (624), 2-Butanone (Methyl Ethyl Ketone) (624), 4-methyl-2-pentanone (624), MTBE (602), m-Xylene & p-Xylene (8021), Naphthalene (602,624), n-Butylbenzene (624), n-Propylbenzene (624), sec-Butylbenzene (624), tert-Butylbenzene (624), trans-1,4-Dichloro-2-butene (624), & Tetrahydrofuran (8260, 624).

The following analytes are Part of ELAP Scope of accreditation but not part of our certification:

Silica (6010), Free Cyanide (4500CN E), Amenable Cyanide (4500DCNG), & Vinyl Acetate (624).

The following Analytes are Part of ELAP Scope of accreditation but not part of our certification for a Non Potable Water Matrix:

Aluminium (200.8), Turbidity (180.1), Methanol (8015D), Dalapon (8151A), 1,2-Dichlorobenzene (601), Acetone (624), MTBE (624), m-Xylene & p-Xylene (602).

The following Analytes are Part of ELAP Scope of accreditation but not part of our certification for a Potable Water Matrix:

Bromide (300), Ethylene Glycol (8015D), Propylene Glycol (8015D).

The following Analyte(s) Part of ELAP Scope of accreditation but not part of our certification for a Solid and Hazardous Waste Matrix:

1,2-Diphenylhydrazine (8270).

The following Analytes are Part of ELAP Scope of accreditation but not part of our certification for an Air Matrix:

1,2-Dichlorobenzene, Carbon tetrachloride, Chlorobenzene, Chloroform, Ethylbenzene, Methylene Chloride, Tetrachloroethene, Toluene, & Trichloroethene.

Definitions and Glossary

Client: Sullivan County BOCES

Job Number:

Abbreviation	These commonly used abbreviations may or may not be present in this report.
%R	Percent Recovery
DL, RA, RE	Indicates a Dilution, Reanalysis or Reextraction.
EPA	United States Environmental Protection Agency
MDL	Method Detection Limit - an estimate of the minimum amount of a substance that an analytical process can reliably detect. A MDL is analyte- and matrix-specific and may be laboratory-dependent.
ND	Not detected at the reporting limit (or MDL if shown).
QC	Quality Control
RL	Reporting Limit - the minimum levels, concentrations, or quantities of a target variable (e.g., target analyte) that can be reported with a specified degree of confidence.
RPD	Relative Percent Difference - a measure of the relative difference between two points.

CUSTOMER: SCBOCES
ADDRESS: 6 Wick Ave
CITY/STATE/ZIP: Little Ferry NY 12754
CONTACT: Sesse Hornell PHONE: 295-4028
PROJECT LOCATION: Talksburg Elementary
PRJ NUMBER/P.O.#:

REPORT TYPE
 STANDARD ISRA
 NJ REG NYASP A B
 OTHER

TURNAROUND
 NORMAL QUICK
 VERBAL

MATRIX: DW=DRINKING WATER S=SOIL O=OIL WW=WASTE WATER
 SL=SLUDGE GW=GROUND WATER

Sample #	DATE	SAMPLING TIME	COMP (Y/N)	GRAB (Y/N)	MATRIX	Client I.D.	Total # of containers	RECEIVED BY	DATE	TIME
1		Kit-SI-1/6/16			DW		40ml Glass HCL Liter Glass HCL 250ml Amber H2SO4 Liter Amber Plain 250ml Plastic HNO3 250ml Plastic NaOH Liter Plastic Liter Plastic H2SO4 250ml Plastic 125ml Plastic Sterile 8oz Glass 2oz Glass 250ml Pl. NaOH/ZnAcc			
2		Kit-SI-2								
3		Kit-SI-1								
4		Kit-SI-2								
5		Kit-SI-1								
6		Kit-SI-2								
7		Kit-SI-1								
8		Kit-SI-2								
9		Kit-SI-1								
10		Kit-SI-2								
11		CAF-SI-1								
12		CAF-SI-2								

REPORT # (Lab Use Only): 105379

NOTE: SAMPLE TEMPERATURE UPON RECEIPT MUST BE 4°C ± 0.5°C
 SAMPLE TEMP: 22.7 °C T

SAMPLE REC'D. ON ICE Y N

PH CHECK Y N

CHLORINE RESIDUAL Y N

NY PUBLIC WATER SUPPLIES

SOURCE ID _____
 ELRP TYPE _____
 FEDERAL ID _____

ANALYSIS REQUESTED: Lead

SAMPLES SUBMITTED FOR ANALYSIS WILL BE SUBJECT TO THE ENVIROTEST TERMS AND CONDITIONS OF SALE (SHORT FORM) UNLESS ALTERNATE TERMS ARE AGREED IN WRITING.

SAMPLED BY: Sesse Hornell COMPANY: SCBOCES DATE: 6/7/16 TIME: 0900
 RELINQUISHED BY: [Signature] COMPANY: SCBOCES DATE: 6/7/16 TIME: 1020
 RECEIVED BY: [Signature] COMPANY: ETR DATE: 6-7-16 TIME: 1020

ROBERT J. WISNETTE CSARH * AS per Sesse - analyze both bottles * both say -1 1/3

105379

OE 6/7/16

Sample #	Date	Time	Matrix	Client ID	250 ml HNO3	Analysis Requested
CAF-B-S1	6/7/16	0900 8:00	DW			Lead
CAF-B-S2						
CAF-G-S1						
CAF-G-S2						
13 CAF-B-S1-1						
14 CAF-B-S1-2						
15 CAF-G-S1-1						
16 CAF-G-S1-2						
17 CAF-D1-1						
18 CAF-D1-2						
19 CAF-D2-1						
20 CAF-D2-2						
21 BAN-S1-1						
22 BAN-S1-2						
23 TL-S1-1						
24 TL-S1-2						
25 TEC-S1-1						
26 TEC-S1-2						
27 BG-S1-1						
28 BG-S1-2						
29 BG-D1-1						
30 BG-D1-2						
31 TEC-D1-1						
32 TEC-D1-2						
33 PEO-S1-1						
34 PEO-S1-2						
35 CUS-S1-1						
36 CUS-S1-2						
37 PD-S1-1						
38 PD-S1-2						
39 NO-S1-1						
40 NO-S1-2						
41 NO-S2-1						
42 NO-S2-2						
43 001-S1-1						
44 001-S1-2						
45 001T-S1-1						
46 001T-S1-2						
47 001-D1-1						
48 001-D1-2						
49 001G-S1-1						
50 001G-S1-2						
001B-S1-1						
001B-S1-2						

2/3

105379

Sample #	Date	Time	Matrix	Client ID	250 ml HNO3	Analysis Requested
51	6/7/16	9:00	DW			Lead
52						
* 53						
* 54						
55						
56						
57						
58						
59						
60						
61						
62						
63						
64						
65						
66						
67						
68						
69						
70						
71						
72						
73						
74						
75						
76						
77						
78						
79						
80						
81						
82						
83						
84						
85						
86						
87						
88						
89						
90						
91						
92						
93						
94		9:00				
95						
96						
97						
98						

3/3

LOGIN SAMPLE RECEIPT CHECK LIST

Client: Sullivan County BOCES

Job Number: 420-105379-1

Login Number: 105379

Question	T/F/NA	Comment
Samples were collected by ETL employee as per SOP-SAM-1	NA	
The cooler's custody seal, if present, is intact.	NA	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	False	
Cooler Temperature is recorded.	True	22.7 C
Cooler Temp. is within method specified range.(0-6 C PW, 0-8 C NPW, or BAC <10 C	False	
If false, was sample received on ice within 6 hours of collection.	True	
Based on above criteria cooler temperature is acceptable.	True	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the sample IDs on the containers and the COC.	True	
Samples are received within Holding Time.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses, incl. any requested MS/MSDs	True	
VOA sample vials do not have headspace or bubble is <6mm (1/4") in diameter.	NA	
If necessary, staff have been informed of any short hold time or quick TAT needs	True	
Multiphasic samples are not present.	True	
Samples do not require splitting or compositing.	True	